




# Applanation Tonometer AT-1

S e r v i n g   y o u r   v i s i o n



## Applanation Tonometer

# AT-1



### Accessories



### Specifications

Model	AT-1				
Model of slit lamp	300XL,700GL 2ZL,4ZL	SM-90 SM-90N	SM-70 SM-70N	SM-30,SM-10 SM-30N,SM-10N 30GL	SM-4,SM-2 SM-4N,SM-2N
Tonometer main unit	S12-17				
Tonometer base	S12-20	—	S12-19	S12-21	S12-15
Measurement method	Depression type (observed with left eye on the slit lamp)				
Depression area	7.354mm <sup>2</sup>				
Diameter of measurement depression area	3.06mm				
Measurement range	0 ~ 80mmHg				
Minimal scale	2mmHg				
Equation for conversion from mmHg to kpa	1mmHg = 101325/760Pa $\approx$ 133.3Pa 100Pa = 0.1kPa				

● Design and specifications are subject to change as improvements are made to the product.

**TAKAGI SEIKO CO.,LTD.**

330-2 IWAFUNE, NAKANO-SHI, NAGANO-KEN, 383-8585, JAPAN  
TEL.+81-269-22-4512 FAX.+81-269-26-6321  
URL:<http://www.takagi-j.com> E-mail:[info@takagi-j.com](mailto:info@takagi-j.com)

**CE**  
**0120**

CB00002 Rev.2 Printed in Japan 2015.08 KY